

<b>Project</b>	CERN-TMS
<b>Date</b>	2007-12-04
<b>Reference</b>	Cern-tms/status-7
<b>Author</b>	Terry Barnaby

## 1. Overview

The full TMS system has been produced and installed at CERN and is under test. Training on the use of the TMS system has been carried out. Some minor issues are being resolved. The FPGA and software source code have been handed to CERN.

## 2. Work Done

The following work has been performed:

1. Added support for second System Controller on live system.
2. Fixed Cycle Parameter generation system to support multiple injection events.
3. The FPGA now loads the PLL\_FREQUENCY on CYCLE\_START as well as after the PLL\_FREQDELAY.
4. Added the ability to view the raw State/Phase table's generated in the TmsControlGui program.
5. Added tmsBackup utility.
6. Fixed a few bugs in the state table generation system.
7. Changed the phase orientation in the phase tables so that a +ve phase value generates a lag in the waveform.
8. Increased the bit width diagnostics Post Trigger delay parameter to 32bits.
9. Fixed a bug where SDRAM data writes could get corrupted when a large amount of data was being read from the system.
10. The Cycle Parameter information has been updated with the additional pllCycleStartFrequency field. This defines the PLL's frequency which is loaded on CYCLE\_START. This setting is intended to set the initial PLL frequency for a calibration period.
11. The Cycle Parameter information has been updated with the bunchMask field. This bitmap defines in which RF bucket's the bunches to be captured lie. bit 0 of this bit mask defines bucket 1 etc.
12. The orientation of the bunchMask, Mean0Mask and Mean1Mask have been reversed so that bucket 1 is bit 0.
13. The system can now handle multiple injection events. The cycle periods are now named CyclePeriodEvent\* rather than CyclePeriodHarmonic to reflect this.
14. The system now has an improved BunchMean system. This calculates the mean of up to 24 individual bunches during data capture as well as the overall mean for all bunches. The additional DataInfo function setting of DataFunctionMean returns data from this system. The top 2MBytes of each channels SDRAM is used for the data storage of the mean values.

15. The BunchMean system now calculates the time for each sample and returns a standard DataValue object with a time field instead of a DataValueMean object that had a numSamples field.
16. The TmsControlGui application has been modified to handle the above changes.

### **3. Meetings at CERN during full system installation**

During the TMS system training visit a number of discussions and meetings were had with various CERN staff. This resulted in a list of minor problems, issues and possible improved features for the TMS system. The list of these is given below:

1. Jeroen has started performing detailed tests of the ADC raw performance. Generally all looks fine. There were some minor ADC issues on three of the PUPE boards. The boards involved are: 104, 108 and 118. Jeroen will send a document covering his findings. The faults are on some of each boards ADC channels and include: 25Mhz noise, clock jitter and what looks like a stuck bit. Jeroen has sent these boards back to Alpha Data and we have installed the three spare boards in their place.
2. There is a slight system design problem in the FPGA firmware/software. The CycleInformation table only stores the timing of the events to nearest millisecond. We need to store the timing more accurately, by storing the actual DataTable address at which the event and associated state change occurred, in the CycleInformation table, so that data read from a cycle period is from just after the event and the data set in SDRAM has the right harmonic number. The DataTable address stored should be that at the start of the orbit, ie synchronised to the local FREF. Note that this will only be possible for CALIBRATION, INJECTION and HCHANGE events where the PLL is locked and the local FREF is valid. For other events the existing ms Timing table system will be used. A small change to the FPGA firmware and associated software is required to fix this.
3. The software's getData() call should set returned parameters to 0 in the event of an error being returned.
4. We will lower the communications timeout to around 200ms so that major faults, like when a Module controller is not responding, returns an appropriate error within a shorter time.
5. We will add the ability to set the TMS system into simulation mode using a single API calls as well as the current system and also provides some means of determining this state using the API. This will help CERN develop the TMS client software while the PS machine is shut down.
6. There may be SDRAM write issues with the current, version 1.2.1, of the FPGA software when setting the simulation test data. We will perform more detailed testing of this.
7. We have seen a Module controller hang very occasionally during or immediately after a FPGA bit file upload. We will investigate this.
8. Currently the TMS system only allows data to be read once a cycle is complete. CERN would like the ability, in the future, to read the data for an individual Cycle Period as soon as it is available. We will look into the issues of achieving this. This work could be done in the future if required.
9. Although the TMS specification did not call for it, it would be useful to have an ejection event. The hardware has three spare digital input lines and one of these could be used for this. The FPGA firmware and software would need modifications to make this work and CERN would need to determine how to provide the ejection signal. This work could be done in the future if required.

#### 4. Work To do

Generally the full TMS system is in a functional state at CERN. There are some small issues to be resolved. The Proton Synchrotron will not be running again until March and so more testing with real BEAM data will need to wait until then. However the client software side can be tested by running the TMS system in simulation mode and tests can be made using the PS's calibration system.

Jeroen will continue with ADC tests and connecting up all of the digital timing and analogue wires and the CERN client side software will be further developed.